



**DEPARTMENT OF THE AIR FORCE**  
HEADQUARTERS 330TH AIRCRAFT SUSTAINMENT WING (AFMC)  
ROBINS AIR FORCE BASE, GEORGIA

26 June 09

MEMORANDUM FOR 77 AESG/TFL  
7980 Lindbergh Landing  
Brooks City-Base, TX 78235-5119

FROM: 330 ACSG/EN  
235 Byron Street, Suite19A  
Robins AFB Ga 31098-16708

SUBJECT: Safe-to-Fly Certification for the Philips Medical Systems Heartstart and Monitor/Defibrillator, Philips MRx, Model M3535A/M3536A.

1. Based on the review of the provided data for this unit, a Safe-to-Fly Certification Letter is granted for the Philips Medical Systems Heartstart and Monitor/Defibrillator Model M3535A/M3536A, for use on C-130E/H aircraft.
2. The Philips MRx is a multi-parameter monitor with twelve lead Electrocardiogram capability, defibrillator, and Automatic External Defibrillator used by medical professionals to provide emergency resuscitation.

