



**DEPARTMENT OF THE AIR FORCE**  
**AEROSPACE SUSTAINMENT DIRECTORATE (AFMC)**  
**ROBINS AIR FORCE BASE GEORGIA**

MEMORANDUM FOR: HQ AMC/TE  
402 Scott Drive, Unit 1 A5  
Scott AFB, IL 62225

FROM: WR-ALC/GRB  
235 Byron Street, Suite 19A  
Robins AFB, GA 31098

SUBJECT: Safe to Fly for C-130E/H/J Operational Assessment of the CCATT system

1. Safe to Fly approval is granted to perform an Operation Assessment (OA) on a C-130E/H/J aircraft using the Deployed Critical Care Air Transport Teams (CCATT). The system consists of an AT&T Tilt 8900 and the Philips Respironics Actiwatch 2. The subject item is the only device authorized for this assessment. The OA shall be carried out in IAW AMC/TE approved test plan.
2. Engineering has determined that there is one technical or safety of flight concern(s) with the use of this device which would require flight restrictions. This determination is based on the following information.
  - a. The device is not to be plugged into the aircraft power. This was not determined to be needed and thus was not tested.

