



DEPARTMENT OF THE AIR FORCE
HEADQUARTERS TACTICAL AIRLIFT DIVISION (AFMC)
ROBINS AIR FORCE BASE, GEORGIA

16 Mar 2011

MEMORANDUM FOR ASC/WNUP Aeromedical Test Lab
2060 Monahan Way
Wright Patterson AFB, OH 45433

FROM: WR-ALC/GRB
235 Byron Street, Suite19A
Robins AFB, Ga 31098-1670

SUBJECT: Safe-to-Fly Certification for the Philips IntelliVue MP2 Patient Monitor

1. Based on the review of the provided data, Safe-to-Fly Certification for the Philips IntelliVue MP2 Patient Monitor is approved for use on C-130E/H aircraft.
2. Testing was completed for the Philips IntelliVue X2. Approval of the Philips IntelliVue MP2 is by similarity from the test results of the Philips IntelliVue X2. The Philips IntelliVue MP2 Patient Monitor is identical to the Philips IntelliVue X2 in every way with one exception. The MP2 does not have the ability to link to another host IntelliVue monitor via companion mode to act as an MMS and to upload data to the host monitor. The Philips IntelliVue MP2 has passed all required testing and is safe for use on C-130 aircraft. Test results indicate the device will not cause adverse effects to the patient, aircrew, or aircraft when used and maintained, by qualified personnel, in accordance with manufacturer guidelines.
3. WR-ALC/GRB shall be notified of any changes to the Philips IntelliVue MP2 for approval prior to use on C-130E/H Aircraft.
4. Questions concerning this approval may be addressed to [REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]