

C-130 Tactical Airlift Division



ROUTING AND TRANSMITTAL SLIP

Date: 8-Feb-2013

TO:

TO:	Office	Name	Action	Digital	Date
1.	WLNEA	[REDACTED]	Coord	[REDACTED]	2/13/13
2.	WLNEA	[REDACTED]	Coord	[REDACTED]	2/13/13
3.	WLN	[REDACTED]	Sign	[REDACTED]	27 Feb 2013
4.					
5.					
6.					
7.	WLNEB	[REDACTED]	Return To		
8.					
9.					
10.					
11.					

REMARKS/SUMMARY

Subject: Safe To Fly, C-130E/H, Philips Medical Systems Heartstart Monitor/Defibrillator Models M3535A/M3536A/M3536M6

1. Recommendation:

Recommend approval. Safe To Fly, C-130E/H, Philips Medical Systems Heartstart Monitor/Defibrillator Models M3535A/M3536A/M3536M6 was certified by the C-130J SPO previously on 8-Aug-2012.

[Privacy Action of 1974 - as amended may apply](#)

This communication could contain personal information which must be protected IAW DoD 5400-11R and is for Official Use Only (FOUO).

FROM: (Name, org symbol) [REDACTED]	PHONE: [REDACTED]
---	-----------------------------



DEPARTMENT OF THE AIR FORCE
AIR FORCE LIFE CYCLE MANAGEMENT CENTER
WRIGHT-PATTERSON AIR FORCE BASE OHIO

27 February 2013

MEMORANDUM FOR AFLCMC/WNUP Aeromedical Test Lab

ATTN: [REDACTED]
2060 Monahan Way Bldg 17
WPAFB, OH 45433

FROM: AFLCMC/WLNE
235 Byron St., Suite 19A
Robins AFB, GA 31098

SUBJECT: Safe-to Fly Certification for the Philips Medical Systems Heartstart
Monitor/Defibrillator Models M3535A, M3536A and M3536M6

1. Based on our review of data for this unit, Safe-to-Fly Certification is granted for use of the Philips Medical Systems Heartstart Monitor/Defibrillator Models M3535A, M3536A and M3536M6 for Aero-Medical Evacuation (AE) missions on the C-130E/H aircraft subject to the stipulations listed below.

- a. When not in use, the device will be stored in either a carrying case or the medic's equipment bag.
- b. When in use, the device will be tied to the litter with litter straps through the device handle and through the two stirrups of the litter as shown in Attachment 1.

2. Points of contact are [REDACTED]

2/27/2013

1 Attachment: Photo of device

Attachment 1

Photo of Philips Medical Systems Heartstart Monitor/Defibrillator Models M3535A, M3536A and M3536M6 properly secured to a litter.

