DEPARTMENT OF THE AIR FORCE
730TH AIRCRAFT SUSTAINMENT GROUP (AFMC)
ROBINS AIR FORCE BASE, GEORGIA 31098

MEMORANDUM FOR AMC/A5QF
ATTN: [redacted]

FROM: 730 ACSG/EN
235 Byron Street, Suite 19A
Robins AFB GA 31098-1670

SUBJECT: Airworthiness Approval for the Nonin Onyx II Pulse Oximeter Model (Report No. FY2007.003, Technical Report of the Nonin Onyx II Pulse Oximeter Model 9550)

1. After reviewing the test data, we have determined the Nonin Onyx II Pulse Oximeter Model 9550 will not fail in a manner that would cause a hazard to the C-5 aircraft.

2. The Nonin Onyx II Pulse Oximeter Model 9550 is suitable for use on the C-5. Final airworthiness for the Nonin Onyx II Pulse Oximeter Model 9550 is granted.

3. The point of contact is [redacted] 730 ACSSS/GFEABB, [redacted] email [redacted]

cc:
HSG/EN/TF
AMC/A5QF